

Certificate for the Proficiency Test Olfactometry 2020

This certificate was issued to confirm the participation of the laboratory:

Ostschweizer Fachhochschule - OST
8640 Rapperswil
Schweiz

with the coding number: 125


in the international proficiency test for olfactometry, held in September 2020. The aim of the test was to determine the precision r and accuracy A_{od} according to EN 13725:2003. Results of the participant are as follows:


Quality Parameter assessed using 1-Butanol in Nitrogen	Results (requirements according to EN 13725)
$A_{od} = 0,176$	complies with requirement: $A \leq 0,217$
$r = 0,278$	complies with requirement: $r \leq 0,477$

The odorant provided was: 1-butanol in nitrogen.

Number of participating laboratories: 36

Kiel, November 5th, 2020


Björn Maxeiner
Director & Head of PT Division


Laura Brosig
Senior Consultant
& Deputy Head of PT Division

